



BIOTEST LABORATORIES PTY LTD

A.B.N. 82 010 924 571
 2 Darnick St, Underwood QLD 4119
 Telephone: (07) 3841 2232
 Email: admin@biotestlab.com.au

CERTIFICATE OF ANALYSIS

Attention:	Quality Assurance Manager LAF Technologies (Melbourne) 12 Royan Place BAYSWATER NORTH VICTORIA 3153	Report Reference	2023-75616
		Issue	1
		Purchase Order:	Not supplied
		Report Date	24 March 2023
		Page Number	1 of 2

Samples received from:	LAF Technologies (Melbourne)
Date samples received:	17 March 2023
Sample Condition on receipt:	Representative 17°C, by infrared thermometer, at 10:00 AM
Sample Storage:	Fridge (5±3°C)
Testing Commenced:	17 March 2023, samples analysed as received.

TEST RESULTS

Biotest Sample No	Sample Description & Batch Number	Microbial Growth TSB 55-60°C/48hrs per item	Growth of <i>Geobacillus stearothermophilus</i> TSB 55-60°C/7days per item
	ARC WA - Batch AS059		
23076/050	C-Control	Detected	Detected
23076/051	# 1	Not Detected	Not Detected
23076/052	# 2	Not Detected	Not Detected
23076/053	# 3	Not Detected	Not Detected
23076/054	# 4	Not Detected	Not Detected
23076/055	# 5	Not Detected	Not Detected
23076/056	# 6	Not Detected	Not Detected
23076/057	# 7	Not Detected	Not Detected
23076/058	# 8	Not Detected	Not Detected
23076/059	# 9	Not Detected	Not Detected
23076/060	# 10	Not Detected	Not Detected
23076/061	# 11	Not Detected	Not Detected
23076/062	# 12	Not Detected	Not Detected
23076/063	# 13	Not Detected	Not Detected
23076/064	# 14	Not Detected	Not Detected
23076/065	# 15	Not Detected	Not Detected
23076/066	# 16	Not Detected	Not Detected

Biotest Method

MTM-06-007-01

MTM-06-007-01

Chemists & Microbiologists

TEST RESULTS continued

Biotest Sample No	Sample Description & Batch Number	Microbial Growth TSB 55-60°C/48hrs per item	Growth of <i>Geobacillus stearothermophilus</i> TSB 55-60°C/7days per item
	ARC WA - Batch AS059		
23076/067	# 17	Not Detected	Not Detected
23076/068	# 18	Not Detected	Not Detected
23076/069	# 19	Not Detected	Not Detected
23076/070	# 20	Not Detected	Not Detected
23076/071	# 21	Not Detected	Not Detected
23076/072	# 22	Not Detected	Not Detected
23076/073	# 23	Not Detected	Not Detected
23076/074	# 24	Not Detected	Not Detected
23076/075	# 25	Not Detected	Not Detected
23076/076	# 26	Not Detected	Not Detected
23076/077	# 27	Not Detected	Not Detected
23076/078	# 28	Not Detected	Not Detected
23076/079	# 29	Not Detected	Not Detected

Biotest Method

MTM-06-007-01

MTM-06-007-01


Key: == Approximate Count
CFU = Colony forming unit

< = Less than
TNTC = Too Numerous to Count

> = Greater than

All samples analysed as received. This report relates specifically to the samples received and replaces all previous revisions and preliminary results. Results relate to the source material only to the extent that the samples are supplied are truly representative of the sample source. All tests were performed at various times between the sample received date and the reported date.

Authorised by:


Ben Yang B.App.Science
Microbiology Manager

TGA Licence No.: MI-04022005-LI-00420-2
APVMA Licence No.: 6036

Final results of the BI's have been analysed, after a full 7 days of incubation. Based on our analysis, none of the biological indicators have recorded any growth. We are pleased to say that a complete 10⁻⁶ has been achieved within the listed, target area(s)