



BIOTEST LABORATORIES PTY LTD

A.B.N. 82 010 924 571
 2 Darnick St, Underwood QLD 4119
 Telephone: (07) 3841 2232
 Email: admin@biotestlab.com.au

CERTIFICATE OF ANALYSIS

Attention:	Quality Assurance Manager LAF Technologies (Melbourne) 12 Royan Place BAYSWATER NORTH VICTORIA 3153	Report Reference	2023-76162
		Issue	1
		Purchase Order:	Not supplied
		Report Date	27 April 2023
		Page Number	1 of 1

Samples received from:	LAF Technologies (Melbourne)
Date samples received:	17 April 2023
Sample Condition on receipt:	Representative 24°C, by infrared thermometer, at 11:15 AM
Sample Storage:	Ambient
Testing Commenced:	17 April 2023, samples analysed as received.

TEST RESULTS

Biotest Sample No	Sample Description & Batch Number	Microbial Growth TSB 55-60°C/48hrs per item	Growth of <i>Geobacillus stearothermophilus</i> TSB 55-60°C/7days per item
23107/116	C- Control, Batch AS059	Detected	Detected
23107/117	Sample # 1, Batch AS059	Not Detected	Not Detected
23107/118	Sample # 2, Batch AS059	Not Detected	Not Detected
23107/119	Sample # 3, Batch AS059	Not Detected	Not Detected
23107/120	Sample # 4, Batch AS059	Not Detected	Not Detected
23107/121	Sample # 5, Batch AS059	Not Detected	Not Detected
23107/122	Sample # 6, Batch AS059	Not Detected	Not Detected
23107/123	Sample # 7, Batch AS059	Not Detected	Not Detected
23107/124	Sample # 8, Batch AS059	Not Detected	Not Detected
23107/125	Sample # 9, Batch AS059	Not Detected	Not Detected

Biotest Method


MTM-06-007-01

MTM-06-007-01

Key: ≈ = Approximate Count < = Less than > = Greater than
 CFU = Colony forming unit TNTC = Too Numerous to Count

All samples analysed as received. This report relates specifically to the samples received and replaces all previous revisions and preliminary results. Results relate to the source material only to the extent that the samples are supplied as truly representative of the sample source. All tests were performed at various times between the sample received date and the reported date.

Authorised by:


Ben Yang B.App.Science
 Microbiology Manager

TGA Licence No.: MI-04022005-LI-00420-2
 APVMA Licence No.: 6036